




EasySense software manual

(PC Windows version)



Data Harvest Group Ltd.
1 Eden Court
Leighton Buzzard
Bedfordshire LU7 4FY
Tel: +44 (0)1525 373666
Fax: +44 (0)1525 851638

e-mail:
sales@data-harvest.co.uk
support@data-harvest.co.uk
Website: www.data-harvest.co.uk

Document No: **DO 135**
Revision 6

Contents

System requirements	4
Installing the EASYSense software	4
Downloading the EASYSense software from our website	4
Connecting an Interface or Data logger	5
Starting the EASYSense program for the first time	5
The Home screen	6
The Meters option	9
Changes	10
Gauge	10
Dial	10
Numeric	10
The Pictogram option	11
The EasyLog option	13
The SnapShot option	14
The Graph option	14
The logging wizard	14
The Scope option	16
The Timing option	17
The Timing Wizard	18
Measurements	21
Apparatus	22
The Counting option	23
The Graph window	25
The Toolbars	25
Recording data	26
Selecting Data	26
Use selection	27
Selection Wizard	27
Previous Selection	27
Analysing captured data	28
Values	28
Interval	28
Difference	29
Gradient	29
Area	29
Show Statistics	30
Displaying data	30
Data Value boxes	30
Axis display	30
Display Options	31
Number of Graphs	31
Graph type	31
Show Grid	31
X-Axis	31
Sensor settings	32
Colour	32
Show or Hide Channels	33
Zoom	33
Auto scale	33
Title	34
The Table display	34

The Comment column	34
Using Tools on the data	34
Notes	34
Test Mode	34
Smoothing	34
Pre-log or Post-log Functions.....	35
Pre-set functions.....	35
Formula functions	39
Best Fit	40
Show Calculation.....	40
Predict	41
Add Text	41
Changing Settings	42
Changing a sensor's range	42
Calibrate Sensor.....	42
Overlay	44
Toolbar	44
Set Language	44
Show Columns	44
Change Units.....	44
Show Data as	44
Deleting data	45
Delete selected.....	45
Edit a value.....	45
Merging data	45
Merge with file	45
Retrieve and Merge.....	45
Screen capture options	45
File menu option	46
Other Home options	47
Open File	47
To open a Worksheet.....	48
To open a Setup	49
To Save your own Setup file.....	49
To associate a Worksheet to your Setup file	50
Setup Remote	50
Retrieve Remote	52
Altering the default configuration for data, setup files and images	53
Adding your own images	53
INDEX	54
EASYSense Software Multi-user Site Licence	57

This manual may be freely copied for use with the **EASYSense** software.

Copyright: all rights reserved.

Data Harvest's policy is to continually improve products and services, so we reserve the right to make changes without notice. It is acknowledged that there may be errors or omissions in this publication for which responsibility cannot be assumed. No liability will be accepted for loss or damage resulting from use of information contained in this manual or from uses as described.

System requirements

Computer:	IBM PC (or compatible)
Operating System:	Windows XP, Windows Vista and Windows 7 (32-bit and 64-bit compatible)
Minimum Software:	Windows Installer 3.1* Microsoft .NET Framework 2* Microsoft Internet Explorer 8.0** Adobe Reader software (V7.0 or above)**
Display:	Minimum 800 x 600 with 256 colours Recommended: 1024 x 768 with high colour - 16-bit

* If Microsoft .NET Framework 2 or Installer 3.1 is not detected during the installation set up process, they will be automatically installed prior to the **EASYSense** application install. This will extend the installation time significantly.

** V8 minimum Internet Explorer is required to view documentation on the EasySense CD. To view and print PDF files you will need Adobe Reader installed.

Installing the EASYSense software

USB users: Make sure the EasySense logger is **not** connected to the USB port prior to, or while, the **EASYSense** software is being installed.

- Log on as ADMINISTRATOR (or with equivalent access rights) on the system.
- Close any open programs – you may need to restart your machine for configuration changes to be made.
- Insert the **EASYSense** CD ROM into the CD Drive. Windows should detect the CD and automatically start up the **EASYSense** auto run program. If auto run fails to start, use Windows Explorer to navigate the CD ROM contents and double click on the **Setup.exe** file.
- Follow the on-screen instructions to complete the installation of **EASYSense** and its associated files* onto the hard disk.
- When the installation is complete, eject the CD ROM and store it safely. Restart your system.

* Complete installs all, whereas a Custom install allows you to deselect the installation of logger manuals or to change their installation path.

A Windows installer package (**.msi**) of the **EASYSense** software is available. The **.msi** package for

- 32 bit systems is supplied in the **i386** directory and for
- 64 bit systems in the **amd64** directory.

(To check your system properties, click on Start, right click on Computer and select Properties).


See [page 53](#) for information on altering the default locations for data and setup files.

Downloading the EASYSense software from our website

The 4500 **EASYSense** software is available as a download from the Data Harvest website (www.data-harvest.co.uk). Save the compressed file to your computer, then 'unzip' it to extract all files before you run the setup.exe or msi.

Note: The method used to unzip a file will depend on the computers operating system and the decompression program on your computer. For example:

1. Download the 4500 EasySense software zip file; select to **Save** it to a temporary directory on your computer.
2. **Right** click on the zipped 4500 directory and select **Extract All** (or Extract or Unzip). Select the extract destination as the temporary directory. When the 4500 directory has decompressed the folder will appear minus its 'zip'.
3. From the unzipped folder either run the Setup.exe or select the appropriate Windows installer package (**.msi**).



Connecting an Interface or Data logger

- **USB connection**

The first time an **EASYSense** unit (e.g. Vu, VISION, 1Link, 3Link, Q3+, Q5+, Q3, Q5, Qt Timer, QLink, or Q Advanced) is connected to a USB port on the computer, the USB drivers will need to install.

Connect the **EASYSense** unit to the USB port on the computer. Windows will automatically detect a new device and install the drivers.

- **Bluetooth™ connection:**

Use the instructions provided in the **EASYSense** logger's user manual.

- **Serial connection:**

Connect the interface/data logger to the serial port on the computer. Note which serial port you have connected to e.g. COM 1, as this information will be required when you start the **EASYSense** software.

Starting the EASYSense program for the first time



Start the **EASYSense** software e.g. Start ► Programs ► Data Harvest ► and click on **EASYSense Software**.

Select the Language

If a choice of language is available a Set Language window will open.


To choose a different language, select from the drop-down menu and OK. The program will close. Restart to apply your choice.

Select a User level

The first time the **EASYSense** program is opened a 'Select Program Level' window will automatically open. Select a suitable user level.

- Level 1 is aimed at 'start' level (e.g. up to 9 years old)
- Level 2 is aimed at 'mid' level (e.g. 9 – 15 years old)
- Level 3 is aimed at 'exam' level (e.g. 15 years plus)

Note: The program level can be altered by clicking on the Level icon and cycling through the levels (one bar for Level 1, two for Level 2 and 3 for Level 3). The currently selected level will be saved when the **EASYSense** program is closed.



Select an Interface and Connection


The first time **EASYSense** is opened a 'Select Interface' window will automatically open to allow the correct interface and connection to be selected.

(a) Select the correct interface from the drop-down menu.

- **EasySense Range** for an **EASYSense** data logger or interface
- **Flash Logger** for the **EASYSense** Flash Logger (XP only)
- **Sense and Control** for the Sense & Control or Sense & Control EasyLog.
- **Example Data** for stored example data to show how sensor readings are displayed. If this option is selected you will not be able to collect data from your interface or data logger.

(b) Select the appropriate connection i.e. either Serial or USB Port.

- If you are connecting via a **USB** lead, select **USB Port**.
- If you are using a **serial** lead, select **Serial Port** and then the number of the communication port used from the serial port drop-down list e.g. Communications Port (COM1).
- If you are using a **USB/serial adapter**, select **Serial Port** and then the adapter from the serial port drop-down list e.g. AMC232USBDH USB Serial Port (COM4).



- If you are using **Bluetooth**, select **Serial Port** and then the serial port configured for Bluetooth from the serial port drop-down list e.g. Bluetooth Communications Port (COM5).

(c) If these settings will stay the same leave the 'Show at startup' box unticked.

If the box is ticked (enabled) the 'Select Interface' window will automatically open when you next start **EASYSense**, so you can check that the currently selected Interface and method of connection are correct. This is useful if you use different types of interface (e.g. both Sense & Control and **EASYSense** Advanced units) or if the interfaces use different methods of connection (e.g. use both **EASYSense** Advanced units which connect via a serial port and **EASYSense** LINK units which connect via a USB port).

(d) Click on OK. The program will save the current selection.

Note: To change this selection at any time, CANCEL any 'failure to connect' errors, select **Interface** from the **File** menu and alter the selection.


The Home screen




Home is the main navigation screen, which contains the icons for all available options.

The appearance of the Home screen and the complexity of the toolbars and functions in the experiment options are altered by the user level selected.


Note: If the interface/data logger detected cannot use an option, it will appear greyed out e.g. the Setup Remote and Retrieve Remote options will only be available for data loggers that support remote data collection.



Level 1



Level 2



Level 3



Meters

The Meters window gives an introduction to sensors and how they respond to change. Up to four section displays can be open at any one time and they can be any combination of sensors and displays.

See [page 9](#).




Pictogram

The value from the sensors can be displayed as images or an abstract colour block or bar chart. Each image represents a 10th of the range (or span) of the sensor and is also displayed as a number from zero to ten.

Note: Pictogram is only available if the User level is set to Level 1.

See [page 11](#).





EasyLog

The window will open pre-set to record sensor values **continuously** as a line graph until stopped.


See [page 13](#).



Snapshot

The window will open pre-set to record sensor values on demand (**manual sampling**) with a bar graph display.

See [page 14](#).




Graph

Use graph to record sensor values against a **chosen time span** as a line graph display. The window opens with a real time logging wizard from which the time span, interval between samples, start condition and triggers for the recording can be chosen.

Note: Graph is only available for User level 2 or 3.

See [page 14](#).




Scope

A simple oscilloscope type display. The data from a sensor is redrawn continuously on the screen. Ideal for demonstrating fast events where there are rapid changes such as ECG traces, sound waveform, etc.

Note: Scope is only available for User level 3.

See [page 16](#).



Timing

Timing is used to allow students to study Time, Velocity, Acceleration, Momentum and Kinetic Energy relationships using switch-type (digital) sensors such as Light gates.

See [page 17](#).







Counting

Use to count events by using switch-type (digital) sensors. The chart type can be displayed as images or a colour block or bar chart oriented either horizontally or vertically.

Note: Counting is only available if the User level is set to Level 1 or 2.

See [page 23](#).



Open File

Used to load a previously saved .SSL or .SID file. Locate the correct file and click on open.

See [page 47](#).

Open Setup



To select and open an experimental setup file. This will pre-set the software for a specific operation.

See [page 49](#).

Open Worksheet

To select and open a worksheet and its accompanying experimental setup file.


See [page 48](#).



Set up Remote


Used for setting up the options to capture data while disconnected from the computer (remote). These options can include the duration of a recording, the time between each sample and triggers.

See [page 50](#).




Retrieve Remote

Use this option to retrieve a data set stored in your data logger's memory. See [page 52](#).







14.7 The Meters option




When Meters is selected from the Home screen, the window will open with live data being displayed from a sensor in a Numeric window.

Click on the Stop icon  to stop data being recorded.

To begin recording again click on the Start icon. 

- Up to four section windows can be open at any one time and they can be any combination of sensors and displays.
- A section window can be opened either by clicking on the appropriate icon on the Toolbar or from the drop-down Display menu.
- To alter the sensor displayed in a section, **right** click with the mouse in the section window and a **Sensor Selection** window will open. Select the sensor to display in that section from the list.
- Click on the close button  to close a section window.
- A section window can be maximized to fill the whole window by selecting the maximize button,  click again to return to ¼ screen size.
- Click on the Home icon  to return to the Home screen



As soon as you open Meters or you click on the **Start/Stop** icon, the data is recorded 'behind the scenes'. This data can be transferred into the Graph window for display and analysis using the Transfer to Graph icon (or select from the Edit menu). When in Graph click on the Back to Meters icon  to return to the Meters window.

Changes

This is the simplest of all four sections. It displays a block of colour that changes to reflect the value being collected by the sensor.

Colour changes happen within the span of a sensor's limits. The minimum and maximum of these limits can be altered in **Sensor Settings** from the Display menu or Options icon.

For example:


Temperature - begins at light blue (coldest) and then works its way up to become dark red (hottest).

Light - begins at black (no light), works through greys until it becomes white (full light).

pH – simulates the colour change of universal indicator.

Sound and other sensors – begin at dark green (minimum), becomes lighter until works through to bright green (maximum).

Note: To alter the sensor displayed, right click in the section window and select the required sensor from the **Sensor Selection** window.



Gauge

This section will display a simple bar gauge that increases and decreases with the measured value.

The scale on the left-hand side is divided into sections to correspond to the sensor it is representing.

The maximum and minimum values recorded will be displayed on the right-hand axis by two short lines if **Show Min Max** from the Display menu is ticked. Use **Reset Min Max** from Display to reset these values.




Dial

This section displays a traditional dial whose pointer moves to the right as the measured value increases and moves to the left as the value decreases.

The scale around the dial is automatically set by the software to correspond to the sensor it is representing.

Two thin lines will display the maximum and minimum values recorded if **Show Min Max** from the Display menu is ticked. Use **Reset Min Max** to reset these values.



Numeric

This section displays the currently measured value as a number.

The maximum and minimum value recorded will be displayed if **Show Min Max** from the Display menu is ticked. Use **Reset Min Max** to reset these values.

Decimal numbers can be switched off by deselecting **Display Decimals** in **Sensor Settings** from the Display menu or Options icon. Without decimals selected, the reading will be rounded up to a whole number.



Sensors

The sensors are automatically identified when the Meter option is selected. To change sensors while Meters is logging, simply replace the sensor with a different type. Meters will automatically change the sensor settings so that the correct calibration scales are displayed. If not logging then the change will be detected once logging has started.


To alter the sensor displayed in a section, **right** click with the mouse in the section window and a **Sensor Selection** window will open. Select the sensor to display in that section from the list.

The display colour, decimals, name and the minimum and maximum limits of a sensor can be altered either from the **Display** menu or **Options** icon. See [page 32](#).

To change the **range** of a sensor e.g. to alter the range of a Heart Rate sensor from Waveform to Beats per minute select **Sensor Config** from the **Settings** menu (see [page 42](#)).



The Pictogram option



The program will open with values from the sensors displayed as images on the chart. Each image represents an assigned amount of data, a 10th of a total range (or span) of the sensor and displays the number from zero to 10.

The block or bar icons can be used to change the images to abstract colour blocks or a bar chart.


Charts can be oriented either horizontally or vertically .

To capture a live reading from the sensors click on the **Store** icon. The maximum number of stored readings is 10.

Stored data can be sorted either by its sample number or by the maximum or minimum value of each sensor.

Stop logging, select the **Sort Data** icon and choose how you want the data organised.


Sorted by sensors Max to Min





The amount of data assigned to each image or block can be altered by changing the range of the sensor and/or its scale.

For example with a Light sensor:

- If its range is 0 – 1 k Lux (1,000 Lux) then each image will represent $1/10^{\text{th}}$ of this range = 100 Lux. If the range was 0 – 10 k Lux (10,000 Lux) then each image will represent 1,000 Lux.
- If the minimum to maximum scale of the 0 – 1,000 Lux range is reduced to 0 – 500 Lux then each image will represent 50 Lux.

To change the **range** of a sensor either use New  or Sensor Config (Settings menu). Select the sensors name, then Change Range. When the set range window opens select the required range, OK.


To change the **scale** of the sensor select Options , Sensor Scale. Double click in the cell with the Min and Max value and type in the new number, OK.




The **name** for the data channel can be changed in this window by double clicking in the cell with the sensors name and typing in a new name.


The picture **image** can be altered by double clicking in the cell showing the current image and selecting a preferred image, OK.

See [page 53](#) on adding new images to those available.



To reset to the original name or scale select the data channel and select Reset to default Min Max or Reset to default Name.


Select New  to set up a new recording.




The range, picture or name of the sensor can also be changed from this window.

Select the sensor first then the change option.

Select Next to show the orientations and display type of the chart or select Finish.





To display the 'real' sensor value select the **Values** icon . Select a data point (a box will highlight the selection), the value and its unit will be displayed in the information window. Click on the Values icon again to deselect.



To change the background colour of the chart select the **Background** tab from the Options icon and select a new colour.

To label a data point:

- Right click in the chart area and select **Add Text**.
- A dialogue box will open. Type in the required text and OK.
- Move the cursor over the data point you want to label, a highlight box will indicate the point selected and the label will appear as a 'ghost' box.



- Click and the label will appear.

To move, edit or delete a label, right click on the label and select the appropriate function.



Tip: Long labels are best displayed when the chart is displayed horizontally.


The EasyLog option



If **EasyLog** is selected from the Home screen, the graph window will open pre-set to record sensor values **continuously** as a line graph until stopped. This makes it ideal for experiments when the time of the recording is unknown.


EASYSense View, Q3 or 5 users: The New logging wizard will open, deselect the tick next to any sensors from which readings are not required.

Click on the **Start** icon  to begin recording.

Logging will start with a time span of 30 seconds and when it has elapsed the time span will double automatically. This will continue until the recording is stopped by clicking on the Stop icon. 



Click on a sensors name to change its range

To change a sensors range select **New** . Select the sensor's name; when the set range window opens select the required range, OK and then Finish.

Choose EasyLog when an X vs. Y graph is required (to plot the readings of one sensor against another). Alter the X-axis to display a sensor (rather than time) – see [page 31](#).

The SnapShot option




If **SnapShot** is selected from the Home screen, the graph window will open pre-set to record sensor values on demand (**manual sampling**) with a bar graph display.


EASYSense View, Q3 or 5 users: The New logging wizard will open, deselect the tick next to any sensors from which readings are not required.


Click on the Start icon  to begin recording.

Click within the graph area to collect a sample.

Click on the Stop icon  to stop data being recorded.

If you click on the Start icon  again you will be given the option to add the new readings to the end of the existing data (if No is selected the existing data will be deleted).

To change a sensor's range select **New** . Select the sensor's name; when the set range window opens select the required range, OK and then Finish.



The Graph option

 User Level 2 & 3



If **Graph** is selected from the Home screen, the graph window will open pre-set to record sensor values against a **chosen time span** as a line graph display. The window will open with a real time logging wizard from which the time span, interval between samples, start condition and triggers for the recording can be chosen. When your selection has been made, click on Finish.

The logging wizard

Stage 1: Sensor selection

- Deselect the tick next to any sensors from which readings are not required. If sensors have been altered after start up, click on the **Redetect** button to identify the change of sensor.
- To change a sensor's range, click on the sensor's name, when the set range window opens select the required range, OK.
- Click on Next.


Stage 2: Recording duration

- Select the length of time that you want to record from the list.
- Select the interval between samples being taken.

The number of samples will automatically alter when the total recording time and the interval between samples are selected. The range of recording times will vary according to which data logger is used.



If a very short recording and intersample time are selected then the recording will be in FAST mode. In this mode there will be no screen display of the data until the recording ends.




REAL TIME recording will change to FAST if a combination of very short recording and intersample times are selected

Note: If you wish to view fast data as it is being recorded then use Scope – see [page 16](#).


- Click on Next.

Stage 3: Select the Start condition as required.

This is used to set the condition required for the recording to start. The choice of start conditions and triggers available on this page will depend on both the recording time selected and the data logger used.




E.g. start conditions for QAdvanced if a FAST recording time is selected



E.g. start conditions for REAL TIME recording time

None - To begin recording as soon as you click on the Start icon .

Trigger on Level - To begin recording when the selected sensor is 'above', 'below', 'rises above' or 'falls below' a set value.




Above
To start recording when the value from a sensor is above a set level.



Below
To start recording when the value from a sensor is below the set level.




Rises above
Logging will not start until the value from a sensor rises up above the set level.



Falls below
Logging will not start until the value from a sensor falls below the set level.

Trigger on Digital Condition - to begin recording when a digital sensor, connected to input A or B, is On, Off, Rising to On or Falling to Off.

Note: This option is only available with certain loggers when a FAST recording time is selected.



If Falling to Off is selected, logging will start when an object has passed through the light beam


If Rising to On is selected, logging will start when the first part of the object breaks the light beam

Select any **Pre Trigger** requirements.


Note: This option will only become available if a FAST recording time with a trigger on level or digital condition is selected.


Pre-trigger will allow a percentage of samples to be stored before the trigger condition is met and the remainder to be taken afterwards.

For example, if 1,000 readings are to be taken and a 25% pre-trigger is selected, 249 readings will be taken before the trigger condition is met and 750 on and after the trigger.



Note: Sufficient time must be allowed to elapse for the pre-trigger samples to be taken e.g. at a relatively slow sample time of 100 milliseconds with 1,000 samples and a 25% pre-trigger, it will take 25 seconds to get to the pre-trigger position.


- Click on Finish.
- Click on the **Start** icon . Recording will start when the start condition you have selected is met. The recording will stop when the selected time span has passed.

Note: If you want to stop recording before the time span has passed, click on the Stop icon .

If None was the selected trigger then the recording will start as soon as the **Start** icon  is clicked.

If recording REAL TIME, with a trigger on level selected, a 'Waiting for Trigger' message will appear in the status bar (below the graph). When the trigger condition is reached the recording will begin.

If recording FAST with a triggered start, a recording data message will display (if there is sufficient time!) until the trigger condition is met. The data collected will download as soon as the recording has finished.




The Scope option



Scope models many of the functions of a simple oscilloscope. It displays 'real-time' plots (as they occur) from the value of a sensor versus time. Unlike a graph that has a beginning and an end, the scope trace is refreshed or 'redrawn' continuously as long as data is being measured.

 User Level 2 & 3

Make your selection from the control panel on the right hand section of the Scope window and then click on the **Start** icon  to begin recording.

Once data has been recorded, it can be treated in the same manner as normal graph data e.g. the X-axis altered from time Reading number.

Note: Overlay and Pre-Log functions are not available.

Timebase


The duration / intervals that are available will depend on the method of connection to the computer and the type of interface connected.

The fastest speeds (with an interval shorter than 20 ms) require both direct USB communication and an interface capable of fast logging.

For example:

- When an **EASYSense Q Advanced** is connected via Bluetooth the fastest interval between readings available is 20 ms. If connected via the USB port, the fastest interval is 20 μ s.
- When **EASYSense Advanced** is connected via the serial port or a USB serial adapter, the fastest interval between readings will be 20 ms.
- The fastest interval between readings available for **EASYSense Q3, Q3+ or Q5** is 20 ms. For Q5+ the fastest interval will be 1 ms.

The timebase (interval between samples and duration) can be changed during logging. Use Faster to decrease and Slower to increase the time period between samples.



Timebase

Interval: 50 μ s Faster

Duration: 12.8 ms Slower

Trigger Mode

☐ Normal

☒ Auto

☐ Single Shot

Trigger Edge

☒ Positive ☐ Negative

Trigger Channel and Level

Light ▼

Level: 500 lx Higher Lower

Pretrigger Position


☐ 0% ☒ 25%

☐ 50% ☐ 75%

☐ 100%

Duration	5.12 ms	12.8 ms	25.6 ms	51.2 ms	128 ms	256 ms	512 ms	1.28 s	2.56 s
Interval	20 μ s	50 μ s	100 μ s	200 μ s	500 μ s	1 ms	2 ms	5 ms	10 ms
Max. no. of sensors	1	2	4	4	4	4	4	4	4

Duration	Roll mode	Roll mode	Roll mode	Roll mode	Roll mode
Interval	20 ms	50 ms	100 ms	200 ms	500 ms

When the interval between readings is 20 ms or slower, the timebase will be in roll mode. In this mode the graph trace is redrawn continuously. When you click on Stop  the trace displayed will 'freeze' and the data will become visible in the table.

If the interval between readings is **10 ms or faster**, Scope will be in trigger mode.

Trigger Mode

- **Normal** – records a trace for the specified duration each time a specific starting condition (trigger) is met.
- **Auto** - operates in a similar way to Normal, but if after a second the specified start condition is not activated, a trace will be recorded. This feature is useful in testing that the trigger settings for an experiment are correct. The trace will continue to be updated at a frequency of once per second.
- **Single Shot** - records one single trace and then stops recording automatically.

Trigger Edge

- **Positive** - the reading from a channel has to rise above a set value to trigger a recording.
- **Negative** – the reading from a channel has to fall through a set value to trigger a recording.

Trigger Channel and Level

- Select the Channel to be used for the trigger condition from the drop-down menu.
- Click on the Higher or Lower buttons to change the threshold value for the trigger. An orange dotted line on the graph will mark the current level.

Pretrigger Position

- Allows a percentage of samples to be stored before the trigger condition is met and the remainder to be taken afterwards. When logging stops a brown dotted line will mark the pre-trigger position on the trace.


The Timing option

The diagram below shows a typical apparatus set up for use with Timing. This option is used to handle data from switch-type sensors such as Light gates and Timing mats which only have two states, ON or OFF.

Connect the digital sensors being used to the interface. Measurements can be taken from either a single sensor connected to Input A, or from two sensors connected to both Input A and B.

To perform calculations for Speed/Velocity, Acceleration, Distance, Momentum and Kinetic Energy, it is necessary to enter the parameters used in an experiment. For example, to calculate speed from A to B, enter the distance between the two digital sensors connected to A and B.

In this set up Light gates are connected to Input A and B on the interface and are used to monitor the movement of a cart down a track. An interrupt card is fixed to the top of the cart. The card will cut through the Light gate's infrared beam.





When Timing is selected from the Home screen, the graph window will be pre-set to record as a bar graph display against reading number. The window will open with a timing wizard from which the measurements required, apparatus and its parameters, and units for display can be chosen (the options available will depend on the level you have selected to work at). When your selection has been made, click on Finish.

The Timing Wizard

User Level 1 has a simple wizard to record Time or Speed at A, Time or Speed from A to B, or Time from A to A (stopwatch).

Step 1	Step 2	Step 3 b
• Time	at A	Units and decimal places
	from A to B	Units and decimal places
	from A to A (e.g. Stopwatch)	Units and decimal places

Step 1	Step 2	Step 3 a	Step 3 b
• Speed	at A	Enter the length of the single interrupt card in millimetres	Units and decimal places
	from A to B	Enter the distance from the Sensor at A to the Sensor at B in metres	Units and decimal places

Step 1: Select whether you want to record Time or Speed. Select Next.


Step 2: Select whether you want to measure:

Time	Speed
<ul style="list-style-type: none"> • at A • from A to B • from A to A (e.g. Stopwatch) 	<ul style="list-style-type: none"> • at A • from A to B

Select Next.

Step 3: If a Time option was selected then the next page will be units and decimals – see 3b. If Speed was selected, the next page will be for entering the measurement required for the calculation of speed.

3a:



Velocity from A to B - enter the distance between the sensor connected to Input A and the sensor at Input B (in metres). Click on Next.



Speed at A - enter the length of the single interrupt card that will pass through the sensor connected to Input A (in millimetres). Click on Next.

3b: Select the number of decimal places and the units for the display of calculated data from the drop-down menus.

Select Finish.



Click on the **Start** icon. When a signal change from the digital sensor/sensors is received, the timing measurement will automatically be entered into the table and graph. The vertical axis scale automatically adjusts to match the largest value collected.

User Level 2 and 3

The pages in the Wizard will vary according to the choice made in the previous step.

Step 1: Select how you want to record the data. Click on Next.

Step 2: Select whether you want to measure at A, at A or B, from A to B, at A then B, etc. Click on Next.

Step 3: Select the apparatus/interrupter you will be using in the experiment. Click on Next.

Step 4: Enter the measurement of the apparatus/interrupter required by the calculation. Click on Next.

Step 5: Select the number of decimal places and the units for the display of calculated data from the drop-down menus. Click on Next.

Step 6: Select, if required, to display any of the columns of data used in the calculation of the final set of data.

$$\text{Acceleration at A} = \frac{\text{Change in Velocity}}{\text{Time Taken}} = \frac{\text{Speed A2} - \text{Speed A1}}{\text{Time A1 to A2}}$$

Note: The option to show these columns are also available from **Show Columns** in the Settings menu.

Click on Finish.

LEVEL 2 and 3 choices are:

Measure		User Level	Interrupter	Measurement
Time	at A	1, 2, 3	-	-
	from A to B	1, 2, 3	-	-
	from A to A e.g. Stopwatch	1, 2, 3	-	-
	at A or B	2, 3	-	-
	at A then B	2, 3	-	-
	period at A e.g. Pendulum	2, 3	-	-

Speed / Velocity	at A	1, 2, 3	Single interrupt card	- length of card (mm)
		2, 3	Double interrupt card	- segment length (mm)
		2, 3	Picket fence	- fence pitch (mm) - calculation increment
		2, 3	Pulley	- none
	from A to B	1, 2, 3	-	- distance from A to B (m)
	at A or B	2, 3	Single interrupt card	- length of card (mm)
	at A then B	2, 3	Single interrupt card	- length of card (mm)

Distance	at A	2, 3	Picket fence	- fence pitch (mm) - calculation increment
		2, 3	Pulley	-

Acceleration	at A	2, 3	Double interrupt card	- segment length (mm)
			Picket fence	- fence pitch (mm) - calculation increment
			Pulley	- none
	at A or B	2, 3	Double interrupt card	- segment length (mm)
	from A to B	2, 3	Single interrupt card	- length of card (mm)
	at A then B	2, 3	Double interrupt card	- segment length (mm)
	change from A to B	2, 3	Double interrupt card	- segment length (mm)

Momentum and Kinetic Energy	at A	3	Single interrupt card	- length of card (mm) - mass of object passing through A (kg)
		3	Double interrupt card	- segment length (mm) - mass of object passing through A (kg)
		3	Picket fence	- fence pitch (mm) - calculation increment - mass of object passing through A (kg)
		3	Pulley	- mass of object passing through A (kg)
	at A or B	3	Single interrupt card	- length of card (mm) - mass of object passing through A (kg) - mass of object passing through B (kg)
	from A to B	3		- distance from A to B (m) - mass of object passing through A (kg)
	at A then A	3	Single interrupt card	- length of card (mm) - mass of object passing through A (kg)
	change from A to A	3	Single interrupt card	- length of card (mm) - mass of object passing through A (kg)
	at A then B	3	Single interrupt card	- length of card (mm) - mass of object passing through A (kg) - mass of object passing through B (kg)
	change from A to B	3	Single interrupt card	- length of card (mm) - mass of object passing through A (kg) - mass of object passing through B (kg)

Raw Times (interrupt edge times)	at A	3	-	-
	at A or B	3	-	-

Measurements

Measurement at A

Uses one digital sensor connected to Input A. Timing starts when the signal from Input A changes (ON) and stops when the signal reverts back (OFF).

- If Light Gates are used with a single interrupt card, the time measured is for how long the card took to pass through the sensor.
- For a Push Switch it will be the length of time that the switch is kept depressed.
- For Timing mats it will be the length of time someone stands on the mat.

Measurement from A to A


E.g. stopwatch

Uses one digital sensor connected to Input A.

Timing will start when the signal at input A changes (ON) and will continue until the signal at A goes OFF and then ON again.



Period at A



E.g. pendulum

Start the pendulum swinging, when it is moving freely click on the **Start** icon.

Timing will start when the pendulum first passes through the Light Gate. When the pendulum passes back through, it will ignore the second signal change and will stop at the third change (when the pendulum arrives back at the Light Gate). The time shown will be for a complete period.

Measurement at A then A

Uses one digital sensor connected to Input A. Time will be measured for a single interrupt card passing through a Light Gate (A1), and then returning back through (A2). E.g. for rebound collisions.

Change from A to A

Uses one digital sensor connected to Input A. The measurement will be for the change in value from A2 to A1 e.g. for a single interrupt card passing through a Light Gate and returning back through.

Measurement from A to B

Uses two sensors connected to Input A and B. A measurement is taken for the time an object takes to travel from the sensor at Input A to the sensor at Input B. Timing will start when the signal change at Input A is detected and stops when the signal at Input B changes.

Measurement at A then B

Uses two sensors connected to Input A and B. Three measurements are recorded of an object travelling from the sensor at Input A to the sensor at Input B. The measurements are:

1. Time to pass through Sensor A
2. Time to pass through Sensor B
3. Time to travel from A to B

Measurement at A or B

Uses two sensors connected to Input A and B. A measurement will be taken from whichever sensor detects the change in signal level. The measurement will be the time to pass through either sensor A or sensor B.

Change from A to B


Uses two sensors connected to Input A and B. A double interrupt card is used to measure the acceleration at A and then at B. The measurement shown will be the change in acceleration (Acceleration at B – Acceleration at A).

Distance at A

Use the Picket fence or Pulley, with a Light Gate connected to Input A, to measure the distance moved. Distance is calculated using the step size of the Data Harvest Pulley, or the pitch length of the Picket fence. Each time a new gap in the Pulley or the pitch of the Picket fence is reached; the distance moved is known.

Raw Times


Timing is usually started when the solid black part of an interrupt card blocks the infrared beam on a Light Gate and changes the logic level from 0 to 1. When the clear part of the card is reached, the infrared beam is restored and its logic state changes back to 0 and the event time is taken. Raw times show the logic levels as 0 and 1 values with the time of the event.



Apparatus



Single Interrupt card

Measure the solid part of the card that will pass through the Light Gate (mm). **Speed** is calculated using the length of the interrupt card divided by the time taken for the card to pass



Double Interrupt card

The two segments of a double interrupt card must be equal in width. Measure the solid part of the card that will pass through the Light Gate (mm).




E.g. Average acceleration can be measured using a single Light Gate with a double interrupt card. Acceleration is calculated using the length of the segments on the double interrupt card.

$$\text{Acceleration} = \frac{\text{Change in Velocity}}{\text{Time taken}}$$

$$\frac{V_2 \left(\frac{\text{Length of segment}}{\text{Time}} \right) - V_1 \left(\frac{\text{Length of segment}}{\text{Time}} \right)}{\text{Time taken}}$$

Picket fence

The picket fence is a multi-segmented interrupt card so a sequence of measurements will be taken.



Measure from the start of one black edge to the start of the next black edge.

To improve the accuracy of readings, a calculation increment can be applied to average the times over fewer pitches.

Note: A calculation increment (of more than 1) will result in fewer measurements than the number of pitches on a fence might indicate.

Pitch length (l)	Suggested calculation increment
60 mm or more	1
50 mm	2
40 mm	2
30 mm	3
20 mm	4
*10 mm	5 - 8

***Note:** If a picket fence with a pitch length of 10 mm is used with buggy type investigations (which produce slower times than gravity investigations), the times can be averaged over less pitches so the calculation increment could be reduced to 5.


When a Picket fence is used to measure speed or acceleration, the measurements shown are for instantaneous values i.e. it took this amount of time to travel this distance and the acceleration was.

Spoked Pulley

The measurements for the Data Harvest Spoked Pulley are pre-defined.

The spokes on the pulley can easily interrupt the Light Gates beam to create false results. Wait until all is ready before you click on the **Start** icon.


When a Spoked Pulley is used to measure speed or acceleration, the measurements shown are for instantaneous values i.e. at this distance, the speed was.







The Counting option





 User Level 1 and 2



This option is used to count events using switch-type (digital) sensors such as Light gates or Timing mats which only have 2 states ON or OFF.

The chart type can be changed to images  or a colour block  oriented either horizontally  or vertically .



Select the **Sums** icon  to cycle through simple calculations based on the counting data ($A + B$, $A - B$, $B - A$, Difference), which will be displayed in the information window.

Select New  to set up a new recording.

Deselect the tick next to any sensors from which readings are not required.

Select the sensor first then the change option.



Select Next to select the orientations and display type of the chart or select Finish.

To change the **name** of a data channel select the sensor then Change Name. Type in a new name, OK.

To alter the picture **image** first select the sensor then Change Picture. Select the preferred image, OK.


See [page 53](#) on adding images to those available.




To change the **background** colour of the chart select the Options icon and select a new colour.


The Graph window

EasyLog, SnapShot, Graph, Scope and Timing use the graph window to display and analyse readings.



The Toolbars




The Toolbars can be positioned at the top, bottom or left or right side of the graph. To move, position the mouse pointer over dotted line to the left of the toolbar, when it becomes a  symbol, click, hold down and drag to its new position.



With some logging options, the table area will automatically be exposed. If not, select **Show Table** from the **Display** menu. If it is not selected, then the graph will automatically fill the window.

The table can still be exposed in this mode by dragging the window divider to the right i.e. position the mouse pointer over the window divider to the left of the screen, when it becomes a $\leftarrow \rightarrow$ symbol, click, hold down the mouse button, and drag to the right to expose the table area.



Recording data



To begin recording, simply click on the **Start** icon.



During logging the Start/Stop icon will show a red square. Click on this icon to stop data being recorded before the selected duration has passed.

If you want to repeat the last experiment (with all of the settings the same) simply click on the **Start** icon again.



If you want to graph a new set of data, without the previous set of data being erased, select **Overlay** before you click on Start. (Not available in Scope or Timing).



To set up a new investigation e.g. to change the duration of a recording in Graph, or to identify a change of sensors or change a sensor's range, select the **New** experiment wizard.



To go from one option to another e.g. from EasyLog to Snapshot, select the **Home** icon.


Selecting Data

To select **individual** readings: -

Click on the reading in the table or on the graph. The reading will be highlighted in the table and marked with a cursor on the graph.

To **deselect** readings:


Click with the left hand mouse button in the area above the graph.



To select **more than one** reading in a series: -


- Click on the first reading of the series, hold down and drag the mouse to the last reading or
- Click on the first reading of the series, hold down the SHIFT key and click on the last reading.

The selected readings will be highlighted in the table and marked by a box in the graph.



To select readings that are **not in series**:-
Hold down the CTRL key and click on each reading or point in turn.

The selected readings will be highlighted in the table and marked with cursor lines in the graph.



Use selection User Level 2 & 3

To display a selection of the recorded data, select the data required from the table or graph area (so it is highlighted). Click on **Use Selection** from the **Edit** menu and the graph will alter to use the selected data only.


If you save while the selected part of the data is displayed, then only the selected data will be saved, (not the full data set).

Selection Wizard User Level 2 & 3

Selection Wizard from the **Edit** menu can be used to select readings at specified intervals.

Type in the first and last reading of the selection and the interval between the readings (e.g. every 10th reading). Click on OK and the selected readings will be highlighted in the table and graph.

If the 'Use the selection' box is ticked the graph will automatically alter to display only the selected data.



Previous Selection User Level 2 & 3


Use **Previous Selection** from the **Edit** menu to revert back to the previous selection.

Analysing captured data


There are a number of tools built into the graph window to allow the analysis of captured data.

Note: Some of these tools will not be available when the X-axis is a data channel i.e. Interval, Difference, Gradient and Area.


Click on **Analysis** in the menu bar to open the Analysis menu.



Analysis menu
at Users **Level 1**



Analysis menu
at Users **Level 2**



Analysis menu
at Users **Level 3**


Values

Select the **Values** icon and move the mouse pointer into the graph area. A cursor will appear which can be moved around the graph using the mouse.



Position the cursor at the point where information from the graph is needed. The corresponding value will be shown in the data value box and highlighted in the table. The information box will indicate the time of the selected data point.

If the X-axis is a sensor, then cross hair cursor lines indicate what the value from the sensors would be if they crossed at the cursor point.



Click on the Values icon again to deselect.

Interval

Select the **Interval** icon. Click on the first point on the graph, hold down and drag the mouse to the second point



A box will be drawn around the section of graph and highlighted in the table.

The information box will display the time of the first point, the time at the second point and the interval between them.

Click on the Interval icon again to deselect.

Click in the title area (just above the graph area) to deselect any selected readings.



Difference

Select the **Difference** icon. Click on the first reading to be used in the comparison, to give a fixed cursor at that point (shows as a brown or black line in the graph and table). Click on a second reading.




Note: Alternatively click on the first reading and then select the Difference icon.

The values in the data boxes will show the difference between the data at the two points.

While Difference is still selected it is possible to repeatedly select a second reading to compare with the first.

Click on the Difference icon again to deselect.



Gradient

≡ User Level 2 & 3

This feature calculates the rate of change of data in the vicinity of the cursor.



When the graph line is horizontal the gradient is zero and values are not changing. A positive gradient shows that values are increasing – the larger the gradient, the faster the change. The calculation is performed over two points either side of the cursor.

Click on the **Gradient** icon and a vertical cursor with a small line will appear to indicate the slope of the graph for whichever data channel is currently selected on the Y-axis. The gradient value for all data channels will be shown in the data value boxes.



Click on the Gradient icon again to deselect. A manual Best Fit line can be used to calculate the gradient between two points – see [page 40](#).

Area


≡ User Level 3

This feature is used to calculate and display the area under a chosen section of the graph.



The calculation for area is performed on the data from all channels. The units correspond to the product of the Y and X-axis.

Click on the **Area** icon. Click on the first reading, hold down and drag the mouse across the appropriate section of the graph. Shading lines will mark the section under the graph for whichever data channel is currently selected on the Y-axis. The values for area will be shown in the data value boxes.



Click on the Area icon again to deselect. Click in the area above the graph to deselect any selected readings.

Show Statistics




Show Statistics calculates and displays statistical information about the data collected from the data channel currently selected on the Y-Axis.

Select the appropriate data channel on the Y-Axis. Select **Show Statistics** from the **Analysis** menu.

- Max (Maximum) – the highest value reading
- Mean – the arithmetic mean.
- SD (standard deviation) – the scatter of a series of measurements about their mean value
- Min (Minimum) – the lowest value reading

If only a part of the original graph is selected (e.g. by using Use selection) then the statistics will be calculated using the data from only the selected part of the graph.




To switch statistics off, deselect Show Statistics from the Analysis menu.


Displaying data

Data Value boxes

The value recorded by a sensor will be displayed in a data value box. The colour used in a data value box corresponds to the colour of the plotted line and the axis label on the graph. Each data channel is allocated a different colour i.e. channel one is red, channel two is blue, and channel three is green. These colours relate to the channel of data, not the type of sensor or the input number e.g. if Temperature sensors are used connected to Input 2 and 4 on the interface, they will be allocated to the first available data channels, Channel 1 (red), and Channel 2 (blue). See [page 32](#) to change the channel colours.

If 'Example data' is selected from the **Interface** menu, simulated data will be available. The data value box will be labelled as Sensor (%).

The window divider can be used to resize the data value boxes i.e. position the mouse pointer over the window divider under the data boxes, when it becomes a  symbol, click, hold down and drag downwards.




Axis display

Click with the left hand mouse button in the space to the left of the graph area to alter the data channel displayed. Subsequent clicks will cycle the Y-axis through all the data channels open.

Click with the right hand mouse button in this space to open the Sensor Setting tab option (see [page 32](#)).

Click with the left hand mouse button in the space below the graph area to cycle through the options available for whatever is currently displayed on the X-axis.



1. Time - select between:

- Elapsed Time – from zero until that point in seconds [Time (s)],
- Elapsed Time in days, hours, minutes and seconds [Time (ss), (mm:ss), (hh:mm:ss), (d.hh:mm:ss)].
- Absolute Time - the time & date from the start of recording until that point [Absolute Time (hh:mm:ss), (d.hh:mm:ss)].

2. Sensor - cycle through the data channels available.

3. Reading Number - no choice is available


Click with the right hand mouse button in the space below the graph area to open the X-Axis tab options. From this tab select whether you would like Time, Reading Number or Channel (e.g. sensor) displayed on the X-axis.

Display Options



The options to alter the overall graph display are available either from Display in the menu bar or from the Options icon.


Note: Other display choices, e.g. Show or Hide Channels, Zoom, Title and Auto scale, which apply to an individual graph, are available by a **right** click in the graph area. See [page 33](#).



The Display menu in the graph window - Users Level 2 and 3

Number of Graphs

Select the radio button next to the number of graphs that you want displayed in the graph area.




Graph type

Use this tab to customise the appearance of the collected data in the graph e.g. the graph type, thickness of line, point style and grid.

Choose whether you want to display readings as a line or bar graph.

- If **Line Graph** is selected, then individual readings are by default joined by a line. There is a choice of four different thickness of line or no line (none). The readings can be marked by a choice of symbols or remain unmarked (none).
- If **Bar Graph** is selected then readings will be shown as bars of colour



Show Grid

If a grid is selected then a faint grating is marked on the graph. There is a choice between a fine and coarse grid.


X-Axis

From this tab select whether you would like Time, Reading Number or Channel (e.g. sensor) displayed on the X-axis.

Time - the selected time span is used for the X-axis.

Reading Number - the X-axis will be set to the number of single readings taken from a sensor, without any regular time interval between them. This axis is used for Snapshot data.

Channel - the X-axis will be set to use one of the data channels e.g. a sensor. This axis is used for an X vs.Y plot.




Note: To alter the data channel shown on the X-axis, click with the left hand mouse button in the space below the graph area to cycle through the data channels.

Sensor settings

From this tab the name and limits displayed for a data channel or sensor can be altered.

Normally the full scale of a sensor is displayed on an axis e.g. the scale for a Temperature sensor will show a minimum of -30 to a maximum of 110°C. If the scale of an axis is too large, the minimum and maximum limits displayed can be altered e.g. reduced to a minimum of 10 to a maximum of 50°C.

The name used for a data channel can also be changed e.g. from Temperature to Room Temperature.



Name: Double click in the cell with the data channel/sensors name and type in the new name. The altered name will be marked by an asterisk.

Minimum & Maximum: Double click on the cell with the Minimum or Maximum value and type in the new number. The altered number will be marked by an asterisk.

* Room Temperature * 0.0 * 40.0

To reset to the original name or scale, click to select the data channel and select Reset to Default Name or Reset to Default Min/Max.

Click on OK to finish.

Note: If the value from a sensor being recorded rises above the maximum or falls below the minimum limits that have been set, Graph will not be able to accurately display the true value.

Display Decimals - Used to switch decimal numbers on and off from the display. If Display Decimals is not selected, the reading will be rounded up to a whole number.


Note: While in Timing, the number of decimal places displayed can be altered - see Change Units on [page 44](#).

Colour

From this tab the colour used by a data channel can be altered.

Double click on the data channel/sensor name. A colour palette will open. Select the colour required and click on OK.


Click on OK to finish.



Individual graph choices

Some display options i.e. Show or Hide Channels, Zoom, Auto scale and Title, are applied on an individual graph basis rather than to all graphs displayed. These are available by a **right click** in the graph area of the individual graph.

'Right click'
context menu




Show or Hide Channels

The data from a channel or sensor can be hidden from view.

Right click in the graph area and select **Show or Hide Channels**. Deselect the tick next to any sensors or channels you want to hide.

Example:

In this example two graphs are open. Show or Hide Channels has been used to deselect one of the sensors from each graph.



Zoom

An area of the graph can be magnified so that it can be seen in greater detail.


Right click in the graph area and select **Zoom In**.

Position the mouse pointer at the top corner of the area to be magnified. Click, hold down and drag the rectangular box over the area to be magnified. Release the mouse button to show the magnified area.

To expand this magnified area further, right click, select Zoom in and repeat the above.

Right click and select **Zoom Out** to zoom out by one step.

Right click and select **Zoom Off** to return to the original graph.



Auto scale

This option scales the Y-axis by using the values of the data collected. **Right** click in the graph area and select one of the four choices. The graph will automatically be adjusted.

Auto scale graph Min Max – applies the Auto scale to all the channels of data on the graph using the minimum and maximum limits of the data.

Auto scale graph 0 to Max– applies the Auto scale to all the channels of data on the graph using a minimum value of zero with maximum of the maximum limit of the data.

Auto scale channel Min Max – applies the Auto scale to whatever data channel is currently selected on the Y-axis using the minimum and maximum limits of the data.


Auto scale channel 0 to Max - applies the Auto scale to whatever data channel is currently selected on the Y-axis using a minimum value of zero with maximum of the maximum limit of the data.

Deselect to return to the original scale.

Title

To add a title to a graph, either **right** click in the graph area and select **Title** or double click with the left mouse button above the graph area.



Type in the text for the title.



The Table display

Sensors/channels columns can be hidden from view. **Right** click in the table area. Deselect the tick next to any sensors/channels you want to hide.

The width of the columns in the table can be increased or decreased. Position the mouse pointer over the line in the column heading. Click, hold down and drag to adjust the width.



The Comment column


Once some data has been collected, you can add text to the Comment column in the Table. Double click on the relevant cell in the Comment column. A dialogue box will appear for you to type in text or numbers from your keyboard.

To Edit, double-click on the comment to re-open the dialogue box and alter as required.


Using Tools on the data

Click on **Tools** in the menu bar to open the Tools menu.

Note: Some tool options i.e. Predict, Delete Predictions and Add Text, which apply to an individual graph, are available by a **right** click in the graph area. See [page 41](#).



Tools menu at
Users **Level 1**




Tools menu at
Users **Level 2 & 3**

Notes



Select Notes from the toolbar or from the Tools menu to open the notes window, which has basic text editing facilities. The text written in this window is automatically saved in a data file. Use this window to write notes that would be useful when setting up the experiment, information about how the data was collected, or your conclusions to the experiment.



Test Mode


Select **Test Mode** from the **Tools** menu to test that your sensor is operating correctly. The current value from the sensor will appear in its data value box.

The Test option will be active until it is deselected or a recording is started.

Smoothing

Used to 'even out' the data of a selected line by applying a moving average.

Select **Smoothing** from the **Tools** menu. Select the data channel that you want to smooth from the drop-down menu. Click on the > button to smooth or < to reverse the process. The smooth level advances at one unit at a time for the > symbol and 5 units for the >> symbol. Click on OK when the smooth level is correct.



Pre-log or Post-log Functions



User Level 2 & 3

Functions are mathematical operations that can be applied to data. Functions can be applied to data as it is being collected using a **Pre-log function** or after the data has been collected using a **Post-log function**.

For example: Select **Pre-log function** from the **Tools** menu to show both Velocity and Distance data from a Motion sensor as logging progresses. There may be a slight delay (in most cases unseen) between collection of data and presenting the pre-log data on the screen e.g. 3 data points are needed before calculation of velocity can commence (acceleration would need 5 points).

Note: Pre-log functions are not available in Scope.

For example: Select **Post-log function** from the **Tools** menu to create a new set of data from data that has already been collected e.g. calculates Power from Voltage and Current data.


- Select whether you want to use a pre-set function or formula to create the set of data.
 - **Pre-set** functions are described in words to make creating the data set simpler.
 - **Formula** - presents mathematical formula for generating data.

Pre-set functions:

1. The top menu indicates the subject/topic area, click on the down arrow to view the options available and select the subject area from the drop-down list.

The lower menu will indicate the function(s) available within the selected subject area, click on the down arrow to view the options available and then select the functions from the list.

The white panel at the bottom of the window gives information about what the function will be calculating, what additional information will be needed for the calculation, units, etc. Click on Next.




2. Select the data channel(s) to be used as the source. Click on the arrow to reveal the list of sensors connected, it is presented in the order of the connected sensor ports. With Post-log functions the list will include any derived data channels. Click on Next.
3. Enter the details for the new set of data. A default name and units maybe already entered but they can be edited. Alter the number of decimal places to be displayed if required.

Select whether the axis limits of this data should either expand to fit or use manual limits.

Expand to fit - when the recording has finished, the axis will automatically redraw if data has exceeded the limits taken from the original data.

Manual – use the limits of the axis as specified in the minimum and maximum boxes.



Note: The default value is based on the default limits of the first source channel used in the calculation. If the maximum value calculated by the function is less than the source channel limits, apply Autoscale (see [page 33](#)) to find the limits of the calculated data.

Enter any parameters required e.g. type in a multiplier, constant or scaler. Numbers can be negative or positive (make sure there isn't a space between the + or – sign and the number). The action of the parameter is often indicated in the white panel. Click on Finish.

4. If you are using a Pre-log Function, click on the **Start/Stop** icon. Data will be calculated and displayed on the graph and table. Values will not be displayed in the data values boxes during logging.

Pre-set functions

SUBJECT	FUNCTION	INFORMATION
Electricity	Resistance	<p>Calculates resistance from Voltage and Current data. Current sensor values can be in units of A or mA, Voltage in mV or V. Identify the sensor's units and select the correct multiplier (shown in the white information box). For low resistances set decimal places to 3.</p> <p>When Current is A and Volts is V, make Multiplier = 1 When Current is A and Volts is mV, make Multiplier = 0.001 When Current is mA and Volts is V, make Multiplier = 1000 When Current is mA and Volts is mV, make Multiplier = 1</p>
	Power	<p>Calculates power from Voltage and Current data. Current sensor values can be in units of A or mA, Voltage in mV or V. Identify the sensor units and select the correct multiplier (shown in the white information box).</p> <p>When Current is A and Volts is V, make Multiplier = 1 When Current is A and Volts is mV, make Multiplier = 0.001 When Current is mA and Volts is V, make Multiplier = 0.001 When Current is mA and Volts is mV, make Multiplier = 0.000001</p>

SUBJECT	FUNCTION	INFORMATION
Motion	Velocity	<p>Velocity, calculated from position, distance or displacement data (first derivative) as $\frac{dx}{dt}$ where x is the distance data and t is the time data.</p> <p>Enter the units as appropriate e.g. Linear Motion – units (m) – m/s; (cm) – cm/s; (mm) – mm/s; (in) – in/s. Pendulum, units (deg) – deg/s.</p>
	Acceleration from position data or distance data	<p>Acceleration calculated from position or distance data (second derivative) as $\frac{d^2s}{dt^2}$ or $\frac{d^2\theta}{dt^2}$.</p> <p>Enter the units as appropriate e.g. Linear Motion – units (m) – m/s/s; (cm) – cm/s/s; (mm) – mm/s/s; (in) – in/s/s. Pendulum – units Angular acceleration – deg/s/s.</p>
	Acceleration from Velocity data (first derivative)	<p>Acceleration calculated as the first derivative of Velocity data (linear)</p> <p>Enter the units as appropriate e.g. Linear Velocity [m/s] – Linear Acceleration [m/s/s].</p>
	Angular Acceleration from Angular Velocity data	<p>Angular Acceleration calculated from Angular Velocity data (first derivative)</p> <p>Enter the units as appropriate e.g. Angular Velocity (revs/s) – Angular Acceleration, units revs/s/s Angular Velocity (rads/s) – Angular Acceleration, units rads/s/s</p>
	Displacement from a reference point	<p>Displacement = the distance from a reference point. The distance value of the reference point is subtracted from the distance data to give the displacement.</p> <p>E.g. with an oscillating spring, the 'reference point' would be the distance value of the centre of the oscillation. The displacement required is the displacement from the centre of oscillation.</p>
	Pendulum - Linear Displacement	<p>Calculates the linear displacement along the arc of the pendulum oscillation. It is calculated from the midpoint of the oscillation.</p> <p>Assumption; the angles are small (less than 20°) and $\sin R = R$ in radians</p>
	Displacement from Velocity	<p>Calculates the displacement from Velocity data.</p>

Energy	Gravitational Potential Energy (G.P.E.)	<p>G.P.E. = $E_p = \text{mass (kg)} \times \text{acceleration due to gravity (m/s/s)} \times \text{distance (m)}$.</p> <p>Use to calculate the change in potential energy of an object between two points in a gravitational field. E.g the GPE gained when a ball bounces to a certain height above the floor. $\Delta E_p = m \times g \times \Delta h$</p>
	Kinetic Energy	Kinetic Energy = $[0.5 \times \text{mass (kg)}] \times \text{velocity}^2$
	Pendulum – Potential Energy	<p>Calculates the difference in the potential energy of the pendulum bob between the centre of the oscillation and any other position in the oscillation.</p> <p>The angular displacement data must be in radians before this function is used (see the Conversion pre-set functions).</p>
	Spring – Elastic Potential Energy	<p>Calculates the change in the elastic potential energy of the spring as a result of the displacement of the spring/mass system from the midpoint of the oscillation.</p> <p>Calculated as $\frac{1}{2}Fx$ (Joules)</p>

General	Tare	<p>Use this function to subtract an offset from a sensors readings (typically used to zero the value from a sensor). Find the tare value before using this function.</p> <p>Note: Make the value negative or positive as appropriate e.g. $(2) - (2) = 0$, $(-2) - (-2) = 0$.</p>
	Asks for Value (<i>Snapshot Pre-log only</i>)	<p>This pre-log function will only be available in SnapShot mode. Each time you click in the graph area to record a value, a dialogue box will appear for you to type in a number from your keyboard.</p> <p>Note: Enter a Name and Units as appropriate.</p>
	Manual Entry	<p>Use this function to create a column full of zeros so that your own data can be entered via the keyboard.</p> <p>(To enter data, double click in a cell and type in your own data).</p>
	Add a constant	Use this function to add a number (constant) to each data point.
	Subtract a constant	Use this function to subtract a number (constant) from each data point.
	Multiply by a constant	<p>Use this function to multiply each data point by a number (constant).</p> <p>Note: Alter the number of decimal places as appropriate.</p>
	Divide by a constant	<p>Use this function to divide each data point by a number (constant).</p> <p>Note: Alter the number of decimal places as appropriate.</p>
	Moving Average (simple)	This is a smoothing function. Each value is the mean of the previous 'a' number of data points e.g. if $a = 5$, then the averages of data points 1 to 5, 2 to 6, 3 to 7 etc. will be taken.
	Mean (x,a)	The sum of all the values in a set of 'a' data points divided by 'a' e.g. if $a = 5$, an mean will be calculated for each set of 5 data points, 1 to 5, 6 to 10, 11 to 15, etc.
	Average of 2 Channels	<p>Adds and divides the data from two Channels to show the average of both channels.</p> <p>Note: Alter the number of decimal places as appropriate.</p>
	Average of 3 Channels	<p>Adds and divides the data from three Channels to show the average of all three channels.</p> <p>Note: Alter the number of decimal places as appropriate.</p>

Titration	Asks for the Volume added (<i>Snapshot Pre-log only</i>)	This pre-log function will only be available in SnapShot mode. Each time you click in the graph area to record a value from the pH Sensor, a dialogue box will appear for you to type in a number from your keyboard. Enter the TOTAL volume added so far.
	First derivative of pH	Creates a sharp peak to identify the point of most rapid change in pH.
	Convert Drops to Volume	Use this function to convert 'number of drops' to a volume measurement. This function needs a value for the volume of one drop. Divide the volume in cm ³ by the number of drops required to fill this volume e.g. 1 cm ³ = 27 drops, 1/27 = 0.037 cm ³ per drop. Note: Alter the number of decimal places as appropriate.

Spirometer	Spirometer Air Flow to Volume	This function converts Air Flow data from the Spirometer to Volume. Volume units = Litres (l)
	Spirometer Volume drift correction (<i>Post-log only</i>)	This function corrects Volume data (which has already been calculated from the Air Flow data) for drift. Note: Volume data must be available to select. The multiplier and constant values are pre-calculated. Volume units = Litres (l)

Channel	Multiply Channel by Channel	This function multiplies the data of two Channels together
	Divide Channel by Channel	This function divides the data from Channel one by Channel two.
	Subtract Channel from Channel	This function subtracts the data of Channel two from Channel one (Channel one – Channel two).
	Add two Channels together	This function adds the data from two Channels together

Oxygen	% Saturation to mg/l	This function converts Oxygen % saturation data to Oxygen in mg/l. You will need to supply the Air Pressure (default 101.3 kPa) and the Temperature (°C).
---------------	----------------------	--

Conversion	Convert Degrees (angles) to Radians	Converts Degrees to Radians. For Pendulum (±20°) adjust the minimum channel limits to -0.4 and maximum channel limits to 0.4.
	Radians to Degrees	Converts Radians to Degrees
	Celsius to Fahrenheit	Converts Celsius to Fahrenheit
	Fahrenheit to Celsius	Converts Fahrenheit to Celsius
	Celsius to Kelvin	Converts Celsius to Kelvin

Formula functions:

1. Click on the down arrow to view the formulae available and select from the list. The white panel at the bottom of the window will give information about what the function will be calculating.

a	Mean (x,a)	$a/(x^b) + c$	$a \exp(bt) + c$
$ax + by$	ax^2	$a tx + b$	$a \log_{10}(bx) + c$
$ax - by$	$a/x^2 + c$	$a t/x + b$	$a \log_e(bx) + c$
axy	$ax^2 + bx + c$	$a \sin(bx) + c$	$a \log_{10}(bt) + c$
ax/y	$at^2 + bt + c$	$a \cos(bx) + c$	$a \log_e(bt) + c$
$a/x + c$	$a dy/dx$	$a \exp(bx) + c$	
$(x - a)/b$	$a dx/dt$	$a \sin(bt) + c$	
Moving average (simple)	$ax^b + c$	$a \cos(bt) + c$	

Click on Next.

2. Select the data channel(s) to be used as the source of x or y. Click on the arrow to reveal the list of sensors connected (it's presented in the order of the connected sensor ports). With Post-log functions the list will include any derived data channels.

Click on Next.

3. Edit the default name and units for the new set of data. Alter the number of decimal places to be displayed if required.

Select whether the axis limits of this data should either expand to fit or use manual limits.

- **Expand to fit** - when the recording has finished, the axis will automatically redraw if data has exceeded the limits taken from the original data.
- **Manual** – use the limits of the axis as specified in the minimum and maximum boxes.

Note: The default value is based on the default limits of the first source channel used in the calculation. If the maximum value calculated by the function is less than the source channel limits, apply Autoscale (see [page 33](#)) to find the limits of the calculated data.

Enter any parameters required e.g. a and b are constants i.e. a number or scaler, c is an offset or intercept number. Numbers can be negative or positive (make sure there isn't a space between the + or – sign and the number).

Click on Finish.

Notes:

- A negative number multiplied by a positive number creates a negative.
- A negative number multiplied by a negative number creates a positive.
- Any number multiplied by zero = 0.
- To remove part of a function set the constant's (a or b) number to zero. E.g. To multiplying a data set by a constant would require an ax function. The function $ax + by$ can be altered to ax by setting b to 0 (0 multiplied by $y = 0$).
- A number multiplied by 1 gives the same number.
- To leave the value of a data set unchanged by a calculation then the value of a or b should be set as 1. E.g. when adding two data sets there is no function $x + y$ but there is $ax + by$. By setting both a and b to 1 makes the function $ax + by = 1x + 1y = x + y$.

If you are using a Pre-log Function, click on the **Start/Stop** icon. Data will be calculated and displayed on the graph and table. Values will not be displayed in the data values boxes during logging.

Best Fit



User Level 2 & 3


Select **Best Fit** from the Tools menu to calculate and produce a straight line which fits most closely to the collected data. It can be applied to the whole set of data or to a selected part.

Note: If you want to apply an automatic line of best fit to only a part of the collected data, make your selection before choosing Best Fit.

- Select the method of Best Fit

Automatic (computer calculated) - the line will be drawn automatically across the entire range of the selected data.

Manual (drawn freehand) - lets you draw the line of best fit between any two points




- Click on Next.

- Select the Axes for the line of Best Fit

Select Y-axis – select the data channel/sensor to which you want to fit to the line from the drop-down menu.

Select X-axis – the current X-axis will be selected by default. If the X-axis is Channel, select the data channel to be used from the drop-down list.



- If a part of the collected data was selected for an automatic line, choose whether you want to apply the Best Fit line to the selected data only.
- Click on OK.

If **Automatic** was selected, the line will be drawn automatically.

If **Manual** was selected, click at the start point, move the mouse and a 'rubber band' line will follow, click at the end point.

A manual best fit line can be adjusted. Click on the end point of the line, hold down and drag to its new position.

To delete the best fit line, right click on the line and select **Delete**.

Information shown

Automatic: $y = bx + a$, $r^2 =$


Manual: $y = bx + a$

y = the data channel selected

x = as selected (i.e. time, reading or a data channel)

b is the constant of the slope (gradient between two points)

a is the intercept



r^2 is a fraction between 0 and 1. A high r^2 tells you the line came very close to the points of data. When $r^2 = 1$, all points line up exactly on a straight line with no scatter (so knowing x would let you predict y perfectly). An r^2 value of 0 means there is no linear relationship between x and y (so knowing x doesn't not help you predict y).

Note: The r^2 value is only available if an Automatic line is drawn.

Show Calculation




Timing only

Use to show how a result has been calculated for a reading in Timing.

Select a reading from the table, and select **Show Calculation** from the **Tools** menu. The formula and measurements used in the calculation will be shown.

Click on Next to see the calculation for the next reading or Cancel to finish.

Note: A calculation will not be available for data collected using a Picket fence or Pulley.



Individual graph choices

Some tool options i.e. Predict, Delete Predictions and Add Text are applied on an individual graph basis rather than to all graphs displayed. These are available by a **right** click in the graph area of the individual graph.


Predict

Used to anticipate the effect of a change to the measured parameter before data is recorded e.g. moving a Temperature Sensor from hot to cold water. Predict lines can be drawn on a graph to predict where you think the graph line will go.

To make a Prediction:

- **Right** click in the graph area and select **Predict**.
- Click at the start point, move the mouse and a thick black line will appear. Use the line to draw a prediction on the graph. Click with the mouse button at different points on the graph to fix the line in position. Right click to stop and deselect **Predict**.
- Record the data as normal to test the validity of your prediction.

Predict can be reselected to make up to twenty predictions on a time graph or up to ten for an XY graph.



To delete any unwanted prediction lines right click in the graph area and select **Delete Prediction**. Use the forward and backward buttons to cycle through the prediction lines. As each line is selected it will turn red on the graph. Click on the **Delete** button to erase the prediction.



Add Text

Used to annotate the graph with 'text boxes' that can be placed in any position in the graph area. Each text box has a line which can be made to 'point' to a part of the graph to which it refers.

- **Right** click in the graph area and select **Add Text**.
- A dialogue box will open. Type in the required text and OK.

Note: Select Enter on your keyboard to move to a new line.

- Move the cursor to the point on the line that you want the label to refer to and click once.
- A 'ghost' rectangular box will appear, move it to the required position for the label. Click and the label will appear.



The label can be moved at any time - click on the label and drag using the mouse.


To edit or delete a label, right click on the appropriate label and select **Edit** or **Delete**.

Changing Settings

Changing a sensor's range


Note: This option is specifically for use with **Smart Q** sensors and cannot be used with Sense & Control and EASYSense Interfaces.

Some *Smart Q* sensors have multiple ranges e.g. a Light Level Sensor. The simplest way to change a sensor's range to one suitable for an experiment is:


- Select the New recording wizard icon .
 - Click on the sensor's name (it will be listed using its current range).
 - A set sensor range window will open; the range currently selected will be highlighted.
- Note:** If the sensor only has one range this window won't open.
- Select the required range, then OK.
 - Select Finish to exit the wizard.



Click on Sensor's name



Or

- Select **Sensor Config** from the **Settings** menu.
 - Select the sensor requiring configuration from the list. Click on the **Change Range** button.
- Note:** If no choice is available a message to say 'This Sensor does not have multiple ranges' will be displayed.
- Select the required range and click on OK.
 - Close Sensor Config. Select the New recording wizard icon  and then Finish for the change in range to be detected.

The sensor range setting will be retained until changed by the user.

Calibrate Sensor

Note: This option is specifically for use with **Smart Q** sensors and cannot be used with Sense & Control and EASYSense Interfaces.


All *Smart Q* Sensors are supplied calibrated but certain Sensors allow the user to adjust the calibration constants to the conditions of the experiment. This option allows the user to adjust the sensor reading at either two or three points in its range. The adjustments made will affect the whole Sensor range, between and beyond these points. The accuracy of the user calibration will depend upon the number of calibration points used and their spacing.

For example:

The pH Sensor has a default calibration set for operation at a temperature of 25°C. If the samples being measured are at a lower or higher temperature, then this default calibration may be improved by a 'User calibration'. Buffer solutions of a known pH are used to set the values for the calibration points. It is essential that the buffer solutions used are at the same temperature as the samples in the experiment and that the pH value entered is the value of the buffer at that temperature. For example, a buffer solution with a pH of 7.01 at 25°C can change its value to 7.10 at 5°C, or 6.98 at 40°C.

- Select **Sensor Config** from the **Settings** menu.
- Select the Sensor requiring configuration from the list.
- Click on the **Calibrate Sensor** button

Note: If no choice is available a message to say 'This sensor does not have User Calibration' will be displayed.

- If you are using 3 test solutions to set the calibration points, check the 'Use Three point calibration' box is ticked. If only 2 test solutions are available, deselect the tick in the 'Use Three point calibration' box. Only the First and Last calibration points will then be available.
- Type in the test solution value of all the buffers that will be used and click on **Next**.
- Rinse the electrode thoroughly with distilled water.
- Place the Sensor in the test solution with the lowest value e.g. pH 4 buffer. Agitate then leave for a minute until the reading shown in the lower section of the Calibrate sensor window stabilizes, click on the **Next** button – check the ADC value is entered alongside the First Values box. Rinse the electrode thoroughly with distilled water.
- Place the Sensor in the test solution with the mid range value e.g. pH 7 buffer. Agitate then leave for a minute until the reading shown in the lower section of the Calibrate sensor window stabilizes, click on the **Next** button – check the ADC value is entered alongside the Second Values box. Rinse the electrode thoroughly with distilled water.
- Place the Sensor in the test solution with the highest value e.g. pH 10 buffer. Agitate then leave for a minute until the reading shown in the lower section of the Calibrate sensor window stabilizes, click on the **Next** button – check the ADC value is entered alongside the Last Values box.
- A message to say 'The User range has been programmed. Disconnect and reconnect the Sensor from the EASYSense unit' will be displayed. Unplug the pH sensor from the **EASYSense** unit and then plug back in. OK the message.
- Click on **Change Range** button and select **User Cal** as the range and click on OK.
- Close Sensor Config. Click on the New  icon then **Finish** for the change in range to be detected.

The User calibration values will be stored in the *Smart Q* Sensor and will be retained until recalibrated.

The lower section of the Calibrate sensor window displays the current reading from the selected sensor.

- Default calibration – shows the current value recorded by the default calibration stored in the Sensor.
- Current User calibration – shows the current value recorded by using the calibration points last configured by the user.
- Raw ADC – shows the value from the analogue digital converter in the **EASYSense** unit in a range of 0 – 65535. This value is used by the software to calculate the calibration constants.

Logger Information

Click on the Logger information button to check the type and the firmware version number of the **EASYSense** unit being used.

Overlay



Select the **Overlay** icon from the toolbar or **Overlay Mode** from **Settings**, to add a new set of data to the graph without the previous set of data being erased - useful for repeating an experiment to compare data.

Note: This option is not available in Scope or Timing.




Snapshot,
EasyLog
and Graph

Toolbar

These options let you choose whether to see small or large icons, and whether the icons show text.

Smaller icons without text can be useful on low resolution displays, as it gives more space to the main graph area.



Set Language

If a choice of language is available then Set Language will be listed in the Settings menu.

To choose a different language, select from the drop-down menu and OK. The program will close. Restart to apply for your choice.

Show Columns

To show data used to calculate the final result select **Show Columns** from the **Settings** menu in Timing.

Tick the box next to any columns you would like displayed from the list. The columns available will vary depending on the Timing mode selected.



Timing only



Change Units



Timing only

Select **Change Units** from the **Settings** menu to specify the units and number of decimal places displayed (0 to 0.000000) in Timing. Each measurement has its own drop-down menu.

	Units
Time	ms, s
Distance	mm, cm, m
Speed/Velocity	mm/s, cm/s, m/s, mph, km/h
Acceleration	mm/s ² , cm/s ² , m/s ²
Momentum	g m/s, kg m/s
Kinetic Energy	mJ, J

Show Data as



Timing only

Data can be recalculated in another form without having to repeat an experiment using **Show data as** from the **Settings** menu. The Timing Wizard will open, select how you want to show the current data and enter any measurements required for the calculation. E.g. recalculate Distance data collected using a picket fence to show Speed.

Deleting data

Delete selected

Select the data that you want to delete – see [page 26](#). Select **Delete selected** from the **Edit** menu and confirm the deletion of the currently selected data.

If the data is recorded with a constant time interval between readings (Graph and EasyLog), the data will be deleted but the time intervals will remain consistent. If readings are displayed as joined by a line, it will redraw between the data points.

If data is recorded with an inconsistent time point (Snapshot), both the data and reading number will be deleted and the graph will redraw.

Edit a value

Double click on the appropriate reading in the table, edit the value and select Enter on your keyboard.


Delete Data Sets

Use **Delete Data Set** from the **Edit** menu to erase an entire set of data.

Choose the channel of data to erase from the drop down menu. Click on Erase and confirm the permanent deletion of the data set from that channel.


Note: Once a data set has been deleted there is no recover option.

 User Level 2 & 3



Merging data

Merge with file

 User Level 3

Select **Merge with file** from the **File** menu to merge the current data with a saved .SSL file. The data must be recorded with a time base (not suitable for Snapshot data) and the intersample time of the data must be the same for both.

Note: You may find it best to set up a recording using a method in which the intersample time for the recording is specified e.g. Graph rather than using EasyLog if you intend to merge the resulting files. This option is not available at User Level 1 & 2.

Retrieve and Merge

 User Level 3

Lets the current data merge with a data set retrieved from an **EASYSense** data logger. The data must be recorded with a time base (not suitable for Snapshot data) and the intersample time of the data must be the same for both.




Note: You may find it best to set up a recording with a specified intersample time e.g. using Setup Remote, to ensure the interval between samples is the same (rather than EasyLog) if the intention is to merge the resulting data. This option is not available at User Level 1 & 2.

Screen capture options

Item	Available in	Description
Copy Table	EasyLog, Snapshot, Graph, Scope, Timer	Used to copy the current Table onto the clipboard from where it can be pasted into spreadsheet, word processor or desktop publishing applications.
Copy Graph	EasyLog, Snapshot, Graph, Scope, Timer	Used to copy the current Graph onto the clipboard from where it can be pasted into word processor or desktop publishing applications.

Copy Window	Meters, EasyLog, Snapshot, Graph, Scope, Timer	Used to copy the current EASYSense window (as displayed on the screen) onto clipboard where it can be pasted into word processor or desktop publishing applications.
Export Bitmap	Meters, EasyLog, Snapshot, Graph, Scope, Timer	Select Export Bitmap from the File menu to capture and save whatever is shown in the graph and/or table area as a bitmap image using a filename and destination of your choice [default location is to the currently logged on user's My Documents directory\EasySense Files\Data Files]. Saved as .BMP type.

File menu option


Item	Available in	Description
Home 	Meters, EasyLog, Snapshot, Graph, Scope, Timer	To return to the Home screen.
New 	EasyLog, Snapshot, Graph, Scope, Timer	Use to set up a new recording e.g. opens Logging wizard.
Open	EasyLog, Snapshot, Graph, Scope, Timer	Use to load previously saved data file. Locate the correct EasySense file (.SSL) or .SID file and click on Open. [Default location is to the currently logged on user's My Documents directory\EasySense Files\Data Files].
Save	EasyLog, Snapshot, Graph, Scope, Timer	Provided the file has been saved previously, the data will be automatically saved, retaining the name given the first time it was saved. If you Save while only a selected part of the data is being used then only the selected data will be saved.
Save As	EasyLog, Snapshot, Graph, Scope, Timer	A dialogue box will open to allow the captured data to be saved as an EasySense file using a filename and destination of your choice. Saved as .SSL type file. [Default location is to the currently logged on user's My Documents directory\EasySense Files\Data Files]. Note: The Save or Save As options are not available in Meters as the data collected is live, but data recorded 'behind the scenes' can be transferred into the Graph window and then saved.
Save Setup	Meters, EasyLog, Snapshot, Graph, Scope, Timer	The settings for an experiment can be saved as a Setup file. This allows the use of the same settings at a future date without going through the set-up process. <ul style="list-style-type: none"> Set up the parameters for the experiment Select Save Setup from the File menu. Enter a name for the experiment (which will be displayed in the Worksheet/Setup window) and click on OK.  <ul style="list-style-type: none"> Type in a file name for the EASYSense Setup file and Save to the currently logged on user's My Documents directory\EasySense Files\Setup Files as an .SSS type file. <p>See page 49 for further information.</p>

Transfer to Excel (User Level 2 and 3 only)	EasyLog, Snapshot, Graph, Scope, Timer	Transfers the collected data into Microsoft Excel, if it is installed on your computer. Excel will open a new worksheet with column headings and data imported from EASYSense .
Export as a .SID/CSV file (User Level 2 and 3 only)	EasyLog, Snapshot, Graph, Scope, Timer	<p>Use to export (save) the collected data in a .SID or CSV file format.</p> <p>From the dialogue box select the Type as either SID or CSV and enter a file name and destination of your choice.</p> <p>Notes:</p> <ul style="list-style-type: none"> Files saved in SID format contain information about units and measurements so allow free exchange of data between a variety of data logging software systems Files saved in CSV format (comma-separated values) are a simple text format suitable for database/spreadsheet systems e.g. Microsoft Excel.
Interface selection	Meters, EasyLog, Snapshot, Graph, Scope, Timer	<p>Used to select the correct type of interface and the type of connection used.</p> <p>Note: If the 'Example Data' option is selected you will not be able to collect data from your interface or data logger. The example sensors will replicate data from four sensors working at different values.</p> <p>The program will save your selection so it will be automatically configured when EASYSense is next used.</p>

Printing


Page Set-up	Meters, EasyLog, Snapshot, Graph, Scope, Timer	To set up the print output to suit your printer. Use to select whether to print in landscape or portrait format.
Print	Meters only	To print out a copy of the current screen display
Print Graph	EasyLog, Snapshot, Graph, Scope, Timer	To print out a copy of the current graph
Print Table	EasyLog, Snapshot, Graph, Scope, Timer	<p>To print out a copy of the current table.</p> <p>Note: A table can use multi-pages to print. Use Print Preview Table to establish the number of pages before you decide whether to print.</p>
Print Preview	Meters only	Use to view the printout before printing.
Print Preview Graph	EasyLog, Snapshot, Graph, Scope, Timer	Used to view the printout of a graph before printing.
Print Preview Table	EasyLog, Snapshot, Graph, Scope, Timer	Used to view the printout of a table before printing.

Other Home options



Open File

Using **Open** then **File** from the Home screen menu is an alternative way to load a previously saved data file. Locate the correct EasySense (.SSL) or .SID file and click on Open.




Note: The default location is the currently logged on user's My Documents directory\EasySense Files\Data Files.

To open a Worksheet


Use **Open** then **Worksheet** from the Home screen menu to open an experimental worksheet together with the software pre-configured for an investigation. Experimental worksheets can be written by the user, or purchased as pre-prepared files. Some example pre-prepared files are included with this program.

A workroom file has both a setup file and a directory (with the same file name), which contains an html file named index.html.



To open pre-prepared files:

- Open the Data Harvest Investigations directory (1).
- Open the Workroom files directory (2).
- Open the subject directory (3).
- Locate the investigation and double click on its setup file (5).



To open user saved files:

- Open the My Investigations directory (4).
- Locate the investigation and double click on its setup file (5).

The worksheet for the investigation will open.

Set up the apparatus as required.

Click on the Launch icon to open the logging option pre-set to record this investigation.




Note: The Cancel icon can be used to close the worksheet and return to the directory window.



Use the To Experiment or To Worksheet icons to swap between the worksheet and the experiment window.




Note: The Forward and Back icons will remain inactive until a hyperlink in the text is selected.



To open a Setup

Use **Open** then **Setup** from the Home screen menu to open the software pre-configured for an investigation. Some pre-prepared setup files are included with this program.



To open pre-prepared setup files:

- Open the Data Harvest Investigations directory.
- Open the Setup files directory.
- Open the subject directory.
- Locate the investigation and double click on its Setup file. One of the logging options e.g. Graph will open pre-set to record this investigation.

To open user files:

- Open the My Investigations directory or select the Open Setup icon.
- Locate the investigation and double click on its Setup file. One of the logging options e.g. Graph will open pre-set to record this investigation.


Note: A Setup that has an associated worksheet can still be opened using Open Setup – the logging option will open without opening the worksheet.

To Save your own Setup file

- Set up the parameters of an experiment using the appropriate logging option i.e. EasyLog, Meters, Snapshot, Graph, Scope or Timing.

E.g. Select the Users level, recording times, interval between samples, any start conditions or pre-triggers plus any options required for display i.e. show table, number of graphs, line thickness and point style, grid on/off, X-axis, name, minimum and maximum limits, title, predictions, deselected channels, pre-log functions, notes, overlay, etc.

- Select **Save Setup** from the **File** Menu.
- Type in a suitable name for the experiment - this is the name that will be displayed in the My Investigations list. Click on OK.




- The Save **EASYSense** Setup file dialogue box will open. Alter the file name if required and Save to the currently logged on user's My Documents directory\EasySense Files\Setup Files (.SSS type file).

To associate a Worksheet to your Setup file


- When you save the Setup file, make a note of the file name used and the save location.

Note: Default location is the currently logged on user's My Documents\EasySense Files\Setup Files directory.

- Create a new directory with exactly the same name as the Setup file, in the same location.
- Load a word processor application capable of saving HTML format files e.g. Microsoft Office Word, Publisher or FrontPage. Create your worksheet document.



- Save the document as an **index.html** file into the directory that has the same name as the set up file.



Note: Ensure that the file is saved as .html - some versions of Microsoft Word will save web pages as .htm files, which will not be recognised by the setup. If so, type in the missing l (so .htm becomes .html).

Note: Applications used to create HTML files (such as MS Publisher) may want to create more than just the index.html file e.g. pictures to go on the page that have been saved as GIF or JPG's. If this is the case, ensure that all of the files are saved in the same directory.

- Select **Open** then **Worksheet** from the **Home** window in **EASYSense**. Select the appropriate directory and then investigation from the list (using the name given for the investigation during the saving process).
- The HTML worksheet will open. Click on the Launch icon to open the logging option pre-set to record the investigation.

Setup Remote

Select **Remote** then **Setup Remote** to program a data logger to collect data independently from the computer. The collected data is stored in the logger's memory to be retrieved by the computer at a later stage.

Note: The Set up Remote option is only available for data loggers that support remote data collection. If the interface selected does not have this facility, the Remote option will either not appear or will be 'greyed out'.

User Level 1 has a simple two stage logging wizard to set up the available options before capturing data.

Stage 1: Sensor selection

- Deselect the tick next to any sensors from which readings are not required.


To change a sensor's range, click on the sensor's name, when the set range window opens select the required range, OK. If sensors have been altered after start up, click on the **Redetect** button to identify the change of sensor.

- Click on Next.

Stage 2: Recording duration

- Select the length of time that you want to record from the list.

The interval between each sample and the number of samples will automatically alter when the time duration is selected. The range of recording times available will vary according to which data logger is used.



- Click on Finish.
- Click on the Start icon  to begin recording data.

User Level 2 and 3 has a three stage recording wizard.

Stage 1: Sensor selection

- Deselect the tick next to any sensors from which readings are not required. If sensors have been altered after start up, click on the **Redetect** button to identify the change of sensor.

To change a sensor's range, click on the sensor's name, when the set range window opens select the required range, OK.

- Click on Next.

Stage 2: Recording duration

- Select the total recording time from the list.
- Select the interval between samples being taken.

The number of samples will automatically alter when the total recording time and the interval between samples are selected.

The range of recording times will vary according to which data logger is used.

- Click on Next.

Stage 3:

- Select the Start condition as required.


This is used to set the condition required for the recording to start. The choice of start conditions and triggers available on this page will depend on both the recording time selected and the data logger used.

When Enter Pressed - To begin recording as soon as you press Enter on the data logger.

When you click on Finish the settings will be downloaded and a message will appear on the computer to inform that 'The logger is now set up for remote logging'. Click on OK.


The data logger can be taken away from the computer and placed in the recording position. It will not start logging until the Enter (or Go) button on the logger is pressed.

Trigger on Level - To begin recording when the selected sensor is 'above', 'below', 'rises above' or 'falls below' a set value e.g. when the value from a temperature sensor rises above 15°C.




Above

For example: -
To start recording when the value from the temperature sensor is above 15°C.




Below

For example: -
To start recording when the value from the temperature sensor is below 15°C.



Rises above

For example: -
Logging will not start until the temperature value rises up above the set level.




Falls below

For example: -
Logging will not start until the temperature value falls down to below the set level.

Trigger on Digital Condition - to begin recording when a digital sensor, connected to input A or B, is On, Off, is rising to on or falling to off.

E.g. when used with Light Gates



If Falling to Off is selected, logging will start when the object has passed through the beam

If Rising to On is selected, logging will start when the first part of the object breaks the beam

Trigger on Time - to begin recording at the time and date specified.

- Click on the Now button to set the current time and date using your computer's clock.
- Alter Trigger on Time to a time and date in the future:
Click on the hour, minute, or second number and then use the up/down arrows to scroll the number up or down.
Click on the date drop-down arrow and select a future date from the calendar.

- Select any **Pre Trigger** requirements.

Note: This option will only become available a trigger on level or digital condition selected.

Pre-trigger will allow a percentage of samples to be stored before the trigger condition is met and the remainder to be taken afterwards. For example, if 1,000 readings are to be taken and a 25% pre-trigger is selected, 249 readings will be taken before the trigger condition is met and 750 on and after the trigger.

Stage 4:

- Click on Finish. Recording will start when the start condition you have selected is met. Recording will stop when the selected time span has passed.

Retrieve Remote

Select **Remote** then **Retrieve Remote** from the Home page to retrieve remotely captured data. Retrieved data can be used in the same way as real-time data i.e. it can be analysed, saved and printed.

Note: This option will only be available for data loggers that support remote logging operations.

EASYSense VISION users:

- Select Interface from **VISION's** Home screen or its PC Connection menu. VISION's screen will alter to show its USB connected.
- Select **Remote** ► **Retrieve Remote** from the PC's Home screen. A dialog box will open showing a list of the sets of data presently stored in VISION.
- Select the required set of data and click on the **Retrieve** button.

Note: If the 'Erase the Data set from the Logger after Retrieve' box is ticked, the data set selected will be deleted after retrieval.


- When you've finished using VISION as an interface, disconnect VISION from the USB port and VISION will restart.

EASYSense users:

Connect the data logger to the computer. Select **Remote** ► **Retrieve Remote** from the Home screen.


The **Date** and **Time** of recording is set by the data logger. **Type** indicates the method of recording e.g. Analogue, EasyLog, Snapshot, and Timing.

Select the required set of data and click on the **Retrieve** button.

**EASYSense Q 3 or 5 users:**

EasyLog or Remote data: Deselect the tick next to any sensors from which readings are not required.

Timing Data: The Timing Wizard will open. Select how you want to retrieve the collected time data, the type of interrupter that was used and any measurement required for the calculation of the data.

**Sense & Control users:**

- Connect Sense & Control to the computer. Ensure it is switched ON (Manual) and is not logging.
- Select **Retrieve Remote** from the Home screen.

Altering the default configuration for data, setup files and images

The default locations for the local data, setup file or pictogram image directory can be altered by changing the **AppConfig.txt** file.

For **32-bit users** this is located at C:\Program Files\DHG\EasySense\Objects. For **64-bit users** it's at C:\Program Files (x86)\DHG\EasySense\Objects.

The default locations are:

```
DHGSetupDirectory=[ProgramFiles]\dhg\EasySense Setup Files
DHGSetupDirectoryV2=[ProgramFiles]\dhg\EasySense Setup Files V2
LocalSetupDirectory=[MyDocuments]\EasySense Files\Setup Files
LocalDataDirectory=[MyDocuments]\EasySense Files\Data Files
DHGPictogramDirectory=[ProgramFiles]\dhg\Pictogram Images
LocalPictogramDirectory=[MyDocuments]\EasySense Files\Pictogram Images
```

Open the AppConfig.txt file into Notepad. Locate and change to point to the directory you wish to set as the new default. Once the AppConfig.txt file has been altered it should be saved.

If the setup files are to be stored onto a shared area on your network (rather than the local machine) then the AppConfig.txt file must be altered on all computers running **EASYSense**.

Adding your own images

You can add pictogram images, saved in a *.jpg and *.png format to the selection available in either the DHG pictogram directory [default location [ProgramFiles]\dhg\Pictogram Images] or locally [MyDocuments]\EasySense Files\Pictogram Images].

INDEX

Absolute time	30	Difference	24, 29
Acceleration	20	Digital condition	15, 52
Acceleration functions	36	Digital sensor	17
Add function	37, 38	Displacement functions	36
Add text	13, 41	Display	30
Analysing data	28	Display decimals	10, 12, 32
Apparatus	22	Display options	12, 31
Area	29	Distance	20, 22
Asks for Value function	37	Divide function	37, 38
Asks for Volume added	38	Double Interrupt card	22
Auto Scale	33	Drops to volume function	38
Average	30, 34	EasyLog	7, 13
Average functions	37	Edit a value	45
Axis	30, 31, 32, 33	Edit comment	34
Back to Meters	9	Edit data	45
Background	13, 24	Edit menu	45
Bar gauge	10	Edit text	13, 42
Bar graph	11, 31	Elapsed time	30
Best fit	40	Elastic Potential Energy function	37
Bitmap file	46	Electricity functions	36
Block graph	11, 24	Energy function	37
Bluetooth connection	5, 6	Enter data function	37
Calculation	24, 35, 40	Erase data from logger	52
Calibrate sensor	42	Erase data set	45
Celsius to Fahrenheit function	38	Example data	5, 30, 47
Celsius to Kelvin function	38	Excel	47
Change from A to A	21	Export as .SID	47
Change from A to B	22	Export as CSV file	47
Change range	12, 42	Export Bitmap	46
Change units	44	Fahrenheit to Celsius function	38
Changes	10	Fast mode	14, 51
Channel colour	32	File menu	46
Channel name	12, 24, 32	First derivative of pH function	38
Collect data	53	Formula	35
Colour change	10	Frequency - average	37
Colour palette	32	Functions	35
Columns	34, 44	G.P.E function	37
Comment	34	Gauge	10
Continuous recording	13, 16	Gradient	29, 40
Conversion functions	38	Graph	7, 14
Copy Graph	45	Graph type	31
Copy Table	45	Graph window	25
Copy Window	46	Grid	31
Counting	23	Hide sensor	12, 33
CSV files	47	Home screen	6, 46
Customising Table	34	Horizontal axis	30, 31
Data channel	30	Horizontal graph	11, 24
Data entry column function	37	Icons	44
Data value boxes	30	Image	12, 24, 53
Decimal places	19, 44	Index.html	50
Decimals	10, 32	Interface	5, 6, 47
Default calibration	43	Interrupter	22
Degrees to Radians function	38	Interval	28
Delete comment	34	Interval between samples	16, 50, 51
Delete data set	45	Investigation	48
Delete predictions	41	Kinetic Energy	20
Delete selected	45	Kinetic Energy function	37
Delete text	13, 42	Label	13, 41
Deselect channel	33, 34	Launch	48
Deselect data	26	Level	5
Dial	10	Light gates	17

Line graph.....	31	Recording data	9, 26
Line of best fit	40	Recording wizard... 12, 13, 14, 18, 24, 42, 50, 51	
Linear displacement function.....	36	Redetect	11, 14, 50, 51
Logger information.....	43	Remote wizard	50
Logging wizard	12, 13, 14, 18, 24, 42, 50, 51	Reset Min Max	10
Magnify area	33	Resistance function	36
Manual data entry function	37	Retrieve	45
Manual sampling	14	Retrieve and merge.....	45
Max/Min	10, 30, 32	Retrieve remote	8, 52
Maximize/Minimise	9	Retrieving fast data	16
Mean.....	30	Save file.....	46
Measurement at A	21	Save Setup.....	46, 49
Measurement at A or B.....	21	Scale.....	12
Measurement at A then A	21	Scale axis	30, 32, 33
Measurement at A then B	21	Scope	7, 16
Measurement from A to A.....	21	Section window	9
Measurement from A to B.....	21	Selecting data.....	26
Merge data file	45	Selection wizard	27
Meters.....	6, 9	Sensor	11, 12, 14, 17, 50
Momentum.....	20	Sensor axis.....	30, 31
Motion functions	36	Sensor colour	32
Multiply functions	37, 38	Sensor Config.....	42
Name	24, 32	Sensor name	12, 24, 32
New..... 12, 13, 14, 18, 24, 26, 42, 46, 50, 51		Sensor range.....	12, 42
None	15, 51	Sensor selection.....	9, 11, 51
Notes	34	Sensor settings.....	10, 12, 30, 32
Number	10, 32	Sensor value boxes.....	30
Number of graphs.....	31	Serial connection.....	5, 47
Number of samples	50, 51	Settings	42
Numeric	10	Setup Remote	8, 50
Offset function	37	Show calculation	40
Open file	8, 46, 47	Show columns	44
Open Setup	8, 49	Show data as.....	44
Open Worksheet.....	8, 48	Show Grid.....	31
Options	12, 31, 47	Show Min Max.....	10
Oscilloscope	7, 16	Show or Hide channel	33, 34
Overlay	26, 44	Show Statistics	30
Oxygen function.....	38	Show Table	26
Pendulum.....	21	SID files	47
Pendulum functions	36, 37	Single Interrupt card	22
Period at A	21	Small icons	44
Picket fence	22	Smart Q sensors	42
Pictogram.....	11	Smoothing	34
Picture image.....	12, 24, 53	Snapshot	7, 14
Pitch length.....	23	Sort data	11
Point style	31	Speed	18, 20
Post-log functions	35	Spirometer functions	38
Potential energy function	37	Spoked Pulley	23
Power function	36	Spring functions.....	37
Precision	44	Standard deviation	30
Predict.....	41	Start condition	
Pre-log functions.....	35	Digital condition	15, 52
Pre-trigger..... 15, 17, 52		None	15, 51
Previous selection	27	Sensor value.....	15, 17, 51
Printing.....	47	Time and date.....	52
Radians to Degrees function	38	Start logging	9, 26
Rate of change	29, 40	Statistics	30
Raw ADC.....	43	Stopwatch.....	21
Raw Times.....	20, 22	Store	11
Reading number axis.....	30, 31	Subject functions	35, 39
Recalculate data	44	Subtract an offset function	37
Record against chosen time span	14, 51	Subtract function	37, 38

Sums.....	24	Trigger when.....	15, 51
Table.....	26	Units	19, 44
Tare function.....	37	USB connection.....	5, 47
Test Mode.....	34	Use selection	27
Text box	41	User calibration.....	43
Time	18, 19	Users level.....	5
Time axis	31	Values.....	13, 28
Time interval	28	Values on demand	14, 37
Timebase	16, 30	Velocity	18, 20
Timespan	50, 51	Velocity functions.....	36
Timing	7, 17	Vertical axis	30
Timing card	17	Vertical graph	11, 24
Timing wizard.....	18, 19	Volume added function.....	38
Title	34	Worksheet	48, 49, 50
Toolbar.....	25, 44	X vs.Y	13, 31
Tools	34	X-axis.....	30, 31
Transfer to Excel.....	47	Y-axis.....	30, 32, 33
Transfer to Graph	9	Zero column function.....	37
Trigger Channel level.....	17, 51	Zoom	33

EASYSense Software Multi-user Site Licence

Definitions

The following expressions have the meanings given here:

- **'DHG'** means **Data Harvest Group Limited**, being owner of all intellectual property rights in the Software
- **'Documentation'** means both printed and electronic user documentation.
- **'Software'** means the program supplied.

Licence

This software package may be freely used, copied, and disseminated. This includes any associated documentation and user guides.

The software may not be amended, changed or altered in anyway.

Limited Warranty and Disclaimer of Liability

The Software is supplied 'as is'; DHG makes no warranty, whether express or implied, as to the merchant ability of the Software or its fitness for any particular purpose. In no circumstances will DHG be liable for any damage, loss of profits, goodwill or for any indirect or consequential loss arising out of your use of the Software, or inability to use the Software, even if DHG has been advised of the possibility of such loss.

General

These conditions supersede any prior agreement, oral or written, between the licensee and DHG relating to the software.